


Instructions for use

Cortisol Urine ELISA

Please use only the valid version of the Instructions for Use provided with the kit

REF**MS E-5100R**

96

RUO

For research
use only –
Not for use
in diagnostic
procedures

Cortisol Urine ELISA

1. INTENDED PURPOSE

For Research Use.

Cortisol Urine ELISA is a manual device intended for the quantitative determination of free Cortisol in human urine from an adult population.

2. PRINCIPLE

The Cortisol Urine ELISA is a competitive enzyme immunometric assay (ELISA) where cortisol (antigen) in the sample competes with the antigenic cortisol conjugated with horseradish peroxidase (HRP) for binding to the limited number of antibodies anti-cortisol coated on the microplate (solid phase).

After the incubation, the bound/free separation is performed by a simple solid phase washing. Then, the enzyme HRP in the bound fraction reacts with the Substrate (H_2O_2) and the TMB Substrate and develops a blue colour that changes into yellow when the Stop Solution (H_2SO_4) is added. The colour intensity is inversely proportional to the cortisol concentration in the sample.

Cortisol concentration in the sample is calculated through a standard curve.

3. REAGENTS, MATERIALS AND INSTRUMENTATION

3.1 Reagents and materials supplied in the kit

Standards and Controls

Cat. no.	Symbol	Standard	Concentration	Volume/Vial
MS E-5101	STANDARD A	Standard A	0 ng/ml	4 ml
MS E-5102	STANDARD B	Standard B	1 ng/ml	1 ml
MS E-5103	STANDARD C	Standard C	5 ng/ml	1 ml
MS E-5104	STANDARD D	Standard D	30 ng/ml	1 ml
MS E-5105	STANDARD E	Standard E	200 ng/ml	1 ml
MS E-5151	CONTROL 1	Control 1	Concentrations of Controls are indicated on the QC-Report.	1 ml
MS E-5152	CONTROL 2	Control 2		1 ml

Content: ProClin > 0.0015%, BSA 0.01%

MS E-5140 **CONJUGATE** Enzyme Conjugate

Content: Cortisol conjugated with horseradish peroxidase (HRP);

ProClin > 0.0015%, BSA 0.1%

Volume: 1 x 33 ml

MS E-5131 **96** Microtiterwells

Content: 1 breakable microplate; Anti-Cortisol antibody adsorbed on the microplate

MS E-0055 **SUBSTRATE** Substrate Solution

Content: H_2O_2 -TMB, 0.26 g/l (avoid any skin contact);

ProClin < 0.0015%

Volume: 1 x 15 ml

MS E-0080 **STOP-SOLN** Stop Solution

Content: Sulphuric acid, 0.15 mol/l (avoid any skin contact)

Volume: 1 x 15 ml

MS E-0030 **WASH-CONC 10x** Wash Solution – 10x Conc.

Content: Phosphate buffer 0.2 M pH 7.4;

ProClin > 0.0015%

Volume: 1 x 50 ml


3.2 Materials required but not provided

- Distilled water

3.3 Auxiliary materials and instrumentation

- Automatic dispenser
- Precision Pipetting Devices
- Microplate reader (450 nm, 620 – 630 nm)

4. WARNINGS

- This kit is intended for research use.
- Use appropriate personal protective equipment while working with the reagents provided.
-  Material of animal origin used in the preparation of the kit has been obtained from animals certified as healthy and the bovine protein has been obtained from countries not infected by BSE, but these materials should be handled as potentially infectious.
- Some reagents (standards, controls, conjugate and wash solution) contain small amounts of ProClin™ 300 (> 0.0015%, < 0.06%) as preservative. Avoid the contact with skin or mucosa.
- Classification according to Regulation (EC) No. 1272/2008 [CLP]
Skin sensitivity, Category 1



Contains: ProClin 300

Warning

Hazard statements:

H317 – May cause an allergic skin reaction.

Precautionary statements:

P261 – Avoid breathing dust/fume/gas/mist/vapours/spray.

P280 – Wear protective gloves/protective clothing/eye protection/face protection/hearing protection.

P321 – Specific treatment (see supplemental first aid instruction on this label).

P333+P313 – If skin irritation or rash occurs: Get medical advice/attention.

P362+P364 – Take off contaminated clothing and wash it before reuse.

- The TMB Substrate contains an irritant, which may be harmful if inhaled, ingested or absorbed through the skin. To prevent injury, avoid inhalation, ingestion or contact with skin and eyes.
- The Stop Solution consists of a diluted sulphuric acid solution. Sulphuric acid is poisonous and corrosive and can be toxic if ingested. To prevent chemical burns, avoid contact with skin and eyes.
- Avoid the exposure of reagent TMB/H₂O₂ to direct sunlight, metals or oxidants. Do not freeze the solution.
- This method allows the determination of Cortisol from 0.47 ng/ml (LOD) to 200 ng/ml.

5. PRECAUTIONS

- Please adhere strictly to the sequence of pipetting steps provided in this protocol.
The performance data represented here were obtained using specific reagents listed in this Instruction for Use.
- All reagents should be stored refrigerated at 2 °C – 8 °C in their original container. Any exceptions are clearly indicated. The reagents are stable until the expiry date when stored and handled as indicated.
- Allow all kit components and specimens to reach room temperature (22 °C – 28 °C) and mix well prior to use.
- Do not interchange kit components from different lots. The expiry date printed on box and vials labels must be observed. Do not use any kit component beyond their expiry date.
- If you use automated equipment, the user has the responsibility to make sure that the kit has been appropriately validated for its intended use/purpose.
- The incomplete or inaccurate liquid removal from the wells could influence the assay precision and/or increase the background.
To improve the performance of the kit on automatic systems it is recommended to increase the number of washes.
- It is important that the time of reaction in each well is held constant for reproducible results. Pipetting of samples should not extend beyond ten minutes to avoid assay drift. If more than 10 minutes are needed, follow the same order of dispensation. If more than one plate is used, it is recommended to repeat the dose response curve in each plate.
- Addition of the TMB Substrate solution initiates a kinetic reaction, which is terminated by the addition of the Stop Solution. Therefore, the TMB Substrate and the Stop Solution should be added in the same sequence to eliminate any time deviation during the reaction.
- Observe the guidelines for performing quality control in medical laboratories by assaying controls and/or pooled urine.
- Maximum precision is required for reconstitution and dispensation of the reagents.

- Samples microbiologically contaminated, highly lipemic, icteric or haemolysed should not be used in the assay.
- Plate readers measure vertically. Do not touch the bottom of the wells.
- Fresh disposable tips must be used when pipetting assay reagents including samples, standards and controls to mitigate the risk of carryover contamination. Failure to do so may lead to invalid results.

6. REAGENT STORAGE AND STABILITY

Store the kit at 2 – 8 °C in the dark.

- The kit is stable at 2 – 8 °C until the expiry date stated on the external kit label.
- Once opened, the standards are stable at 2 – 8 °C for 6 months.
- Once opened, the conjugate is at 2 – 8 °C for 6 months.
- The diluted wash solution is stable for 30 days at 2 – 8 °C.

Important note: open the bag containing the Coated Microplate only when it is at room temperature and close it immediately after use.

7. SAMPLE COLLECTION AND STORAGE

The assay should be performed using urine samples.

Sample Storage	Duration
-20 °C	< 6 months

8. PROCEDURE

8.1 Preparation of the Standards and Controls

Before use, mix for 5 minutes with a rotating mixer.

The standards are ready for use and have the following concentration:

	Standard A	Standard B	Standard C	Standard D	Standard E
ng/ml	0	1	5	30	200

The controls are ready to use; the concentration is printed on the label.

8.2 Preparation of Conjugate

The conjugate is ready to use. Mix gently, for 5 minutes, on a roller mixer.

8.3 Preparation of Wash Solution

Dilute the content of the vial **WASH-CONC 10x** with distilled water to a final volume of 500 ml prior to use. For smaller volumes respect the 1:10 dilution ratio.

It is possible to observe the presence of crystals within the concentrated wash solution; in this case mix at room temperature until the complete dissolution of crystals. For greater accuracy, dilute the whole bottle of concentrated wash solution to 500 ml, taking care also to transfer crystals completely by rinsing of the bottle, then mix until crystals are completely dissolved.

8.4 Preparation of the Sample

The determination of cortisol can be performed in human urine samples.

Important note: the kit has been designed to be used on untreated urine samples; acidification treatments of the urine that lead the pH to values below 5.0 could interfere with the assay and produce aberrant results.

It is not necessary to dilute the sample. The total volume of urine excreted during a 24-hour should be collected and mixed in a single container.

Store the sample at -20 °C if the determination is not performed on the same day of the sample collection.

Before using, mix gently, for 5 minutes, with a roller mixer.

8.5 Procedure

Allow all reagents to reach room temperature (22 °C – 28 °C) for at least 30 minutes.

At the end of the assay, immediately store the reagents at 2 – 8 °C: avoiding long exposure to room temperature.

Unused coated microwell strips should be released securely in the foil pouch containing desiccant and stored at 2 °C – 8 °C.

To avoid potential microbial and/or chemical contamination, unused reagents should never be transferred into the original vials.

As it is necessary to perform the determination in duplicate in order to improve accuracy of the test results, prepare two wells for each point of the standard curve (Standard A – E), two for each Control, two for each sample, one for Blank.

Reagent	Standard	Sample/Control	Blank
Standard A – E	10 µl		
Sample/Control		10 µl	
Conjugate	300 µl	300 µl	
Incubate 1 h at +37 °C ± 0.5 °C. Remove the contents from each well; wash the wells 3 times with 350 µl of diluted wash solution. Important note: during each washing step, gently shake the plate for 5 seconds and remove excess solution by tapping the inverted plate on an absorbent paper towel. <u>Automatic washer:</u> if you use automated equipment, wash the wells at least 6 times.			
Substrate Solution	100 µl	100 µl	100 µl
Incubate at room temperature (22 °C – 28 °C) for 15 minutes in the dark.			
Stop Solution	100 µl	100 µl	100 µl
Shake gently the microplate. Read the absorbance (E) at 450 nm against a reference wavelength of 620 – 630 nm or against Blank within 5 minutes.			

9. QUALITY CONTROL

Good Laboratory Practice (GLP) requires the use of quality control specimens in each series of assays in order to check the performance of the assay. Controls should be treated as unknown samples, and the results analysed with appropriate statistical methods.

The kit controls provided in the kit should be tested as unknowns and are intended to assist in assessing the validity of results obtained with each assay plate.

The mean concentration of each control level is documented in the QC report included with each kit. These mean concentration levels are determined over several assays which are run in duplicate in multiple locations across each plate.

The manufacturer recommends the users to maintain graphic records of the control values generated with each assay run, including the running means, SDs and %CVs. This information will facilitate the controls trending analysis relating to the performance of current and historical control lots relative to the supplied Quality Control data. The trending will assist in the identification of assays which give control values significantly different from their average range.

When interpreting control data, users should note that this product was designed and developed as a manual product. The range stated on the QC certificate should be appropriate for assays that are performed manually and with strict adherence to the Assay Procedure described above. It is recognised by Quality Control professionals, that as a result of differences in conditions and practices, there will always be variability in the mean values and precision of control measurements between different laboratories⁶.

10. CALCULATION OF RESULTS

A variety of data reduction software packages are available, which may be employed to generate the mean standard curve and to calculate the mean concentrations of unknown samples and controls. A 4-parameter logistic (4PL) curve fit, including Standard A is required. Other curve fitting algorithms are not recommended.

Alternatively, a standard curve may be prepared on semi-log graph paper by plotting mean absorbance on the Y-axis against concentration of analyte on the X-axis. Standard A should be included in the standard curve. Read the mean absorbance value of each unknown sample off the curve.

In order for the assay results to be considered valid the kit standards and control must fall within the specifications detailed in the lot specific certificate of analysis.

If a control is out of its specified range, the associated test results are invalid and samples must be retested.

To calculate the cortisol concentration in urine, calculate as above and correct for total volume of volume of urine collected in 24 hours:

$$\text{ng/ml} \times \text{vol (ml) urine 24h/1000} = \mu\text{g Cortisol/24h}$$

11. EXPECTED VALUES

To determine the normal range for urine samples, 128 apparently healthy male and female adults were tested:

	n	Normal range urine (24 h)
Adults	128	1.5 – 63 µg/24h

12. PERFORMANCE AND CHARACTERISTICS

Representative performance data are shown. Results obtained at individual laboratories may vary.

12.1 Analytical sensitivity

Analytical sensitivity was investigated through the LOB (white limit), the LOD (detection limit), the LOQ (quantification limit) and the anal sensitivity (A.S.).

The following table shows the criteria of the study and the results obtained.

	Criteria	Result (ng/ml)
LoB	60 replicates of Standard A, used as "Blank", have been investigated in 5 different sessions over 3 days.	0.28
LoD	6 urine samples with low cortisol concentration have been investigate over 10 assays in duplicate, performed in 5 days.	0.47
LoQ	6 urine samples with low cortisol concentration have been investigate over 10 assays in duplicate, performed in 5 days.	0.56
A.S.	20 replicates of Standard A and 5 replicates of Standard B have been assayed. A.S. has been calculated by linear regression.	0.22

12.2 Precision and reproducibility (complex precision)

Precision and reproducibility have been assessed through 6 different urine samples with different concentration of Cortisol.

The table below shows the Within Run and Total CV%.

Sample	n	Mean (ng/ml)	Within Run CV%	Total CV%
PS2	20	112.141	6.6%	12%
PS4	20	64.563	8.1%	12%
CT High	20	50.577	7.3%	11%
PS5	20	25.878	7.6%	10%
PS6	20	9.269	7.6%	11%
CT Low	20	3.438	7.0%	9%

12.3 Analytical specificity

12.3.1 Interfering substances

Interference for Albumin, Acetylsalicylic Acid, Ibuprofen and Ascorbic Acid were studied by adding the interfering substance to the urine sample with a low and high Cortisol concentration, and by comparing its concentration to the unspiked sample.

The interference has been evaluated as "significant" if it causes a concentration bias > 10% between spiked and unspiked sample.

The following table shows the results obtained:

Substance	Concentration	Interference
Albumin	5 mg/dl	No
Acetylsalicylic acid	3.62 mmol/l	No
Ibuprofen	2.42 mmol/l	No
Ascorbic Acid	5 mg/l	No

Conclusion: no interference has been found for Albumin, Acetylsalicylic Acid, Ibuprofen and Ascorbic Acid.

12.3.2 Cross-reactivity

The cross-reaction of the antibody calculated at 50% according to Abraham is shown in the table:

Reagent	Cross-reactivity
Cortisol	100%
Prednisolone	46.2%
11-Deoxycortisol	4%
Cortisone	3.69%
Prednisone	3.10%
11 α OH Progesterone	1%
Progesterone	< 0.1%
Aldosterone	< 0.1%
Pregnenolone	< 0.1%
17 β Estradiol	< 0.1%
Estrone 3-solfato	< 0.1%
Estriol	< 0.1%

Reagent	Cross-reactivity
Testosterone	< 0.1%
Spironolactone	< 0.1%
DHEA	< 0.1%
DHEA-S	< 0.1%
Androstenedione	< 0.1%
Androsterone	< 0.1%
DHT	< 0.1%
Danazol	< 0.1%
Cholesterol	< 0.1%
Dexamethasone	< 0.1%

12.4 Correlation

137 urine samples were tested with the Cortisol Urine ELISA kit and with a LC-MS method (reference).
The linear regression curve is:

n	Slope	Intercept (ng/ml)	R ²
137	1.008	-0.5019	0.83

13. WASTE MANAGEMENT













Reagents must be disposed of in accordance with local regulations.

All materials that have come into contact with samples and reagents must be disposed of in accordance with country, state and local regulations.

14. BIBLIOGRAPHY

1. Foster, L. B. and Dunn, R. T. Clin. Chem. 20/3 365 (1974)
2. De Lacerda, et al J. Clin. Endocr. and Metab. 36,227 (1973)
3. Roller, E., et al Clin. Chim. Acta 66 319 (1976)
4. Kobayashi, Y., et al Steroids, 32 no 1 (1978)
5. Akarawa, et al Anal. Biochem. 97 248 (1979)
6. Basic QC Practices On-line Course; <http://www.Westgard.com>.

Symbols:

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Use-by date		Batch code		
	Consult instructions for use		Content		
	Caution		Catalogue number		Distributor
	Date of manufacture				For research use only!